

# pH Meter



## Specifications:

	pH	mV	Temperature		
Measurement Range	: 0.00 - 14.00	± 1999	-5 ... 105 °C		
Resolution	: 0.01/0.001	1/0.1	0.1 °C		
Relative Accuracy	: ± 0.01/0.004	± 1/ ±0.4	± 0.5 °C		
pH-calibration	: 1, 2, or 3 points				
Isopotential Point	: 7.00 pH				
Calibration Buffer	: B1: (25 °C)	7.00	4.00	10.01	1.68
	B2: (25 °C)	7.00	4.01	9.21	2.00 11.00
	B3: (20 °C)	7.00	4.00	9.00	2.00 12.00
	B4: (25 °C)	6.86	4.01	9.18	1.68
	B5: User-defined				